Right-Hand Veterinary Necropsy Sink Station

MODEL USPE-PET-NC-RH | PLANNING & CARE MANUAL

Form USPE-PET-NC-RH-PCM-001 · Rev. A · October 28, 2026

Place the sink on the right of the veterinary necropsy workstation to match the room layout. Review the approach to the sink, instrument drawer and adjoining casework without assuming transfer-equipment compatibility.

Document scope

USPE PET veterinary equipment planning and care reference. The equipment platform is unchanged; the veterinary application requires review. Not a released operating or installation manual.

Plan before you connect

Qualified personnel must verify dimensions, support, service access, utilities, backflow protection and electrical protection. Isolate utilities before maintenance. Do not commission damaged equipment or equipment with leaks, loose supports or electrical defects.

Veterinary application, specimen dimensions and load, component compatibility, drainage, electrical supply and any ventilation requirement must be confirmed in the written quote and approved project documents. This station is not represented as a biosafety cabinet, containment system or clinical fluid-administration device.

USPE PET supplies veterinary necropsy stations, animal-science equipment and coordinated laboratory casework. A division of American Mortuary Coolers.

Equipment and configuration reference

- Right-hand sink configuration for layout-specific prep rooms
- 304 stainless steel construction for sanitation and corrosion resistance
- Hydro aspirator with vacuum breaker (standard)
- Dissection rinse valve and control valve
- Flexible spray hose for cleanup and rinsing
- Waterproof GFCI receptacle for wet work areas
- Large instrument drawer and perforated specimen shelf
- Custom specifications available (layout, plumbing, ventilation, size)
- Any separate table pictured alongside the station is not included unless itemized in the quote.

Listed specifications

Material	304 stainless steel
Included utilities	Hydro aspirator, vacuum breaker, spray hose, dissection rinse valve, GFCI receptacle
Storage	Large instrument drawer and perforated specimen shelf

Exact quote configurations

Select the catalog code and listed configuration. A selection is a quotation request, not a purchased item or an installation approval.

USPE PET catalog code	Documented configuration
USPE-PET-NC-RH-01	Standard Drain
USPE-PET-NC-RH-02	2□ Drain Upgrade

Casework and room handoff

Schedule the station with its base, wall and tall cabinets. Mark quantity, room zone, cutouts, handedness, height, support, projections and service clearances. Confirm finishes, glass and hardware in the quote.

Casework material, glass and door-style form

Stainless steel laboratory casework is our specialty. Choose standard stainless steel, powder-coated stainless steel or powder-coated steel. The 100% American-made casework message does not establish the origin of every purchased fixture or accessory.

Cabinet / room zone / quantity: _____
Material / grade / gauge: _____
Body color & tier / front color & tier: _____
Glass appearance: clear / frosted / grey-tinted / bronze-tinted / not applicable
Door style: solid hinged / glazed hinged / sliding / open storage
Handing / frame / glass thickness / safety-glazing requirement: _____
Handle / hinge / sample / drawing revision: _____

Glass appearance is a request, not a stocked glazing specification or separate SKU. Confirm tempered or laminated requirements, thickness, edge treatment, hardware and cabinet fit in the written quote. No impact, fire, containment or cleanroom rating is implied.

Receiving and commissioning

Check condition, model and options against the approved schedule. Report damage or missing components before use. Obtain component instructions, complete utility checks and document commissioning. A photograph does not establish capacity, containment or waste-disposal approval.

Care and service

Follow component instructions and the facility's approved cleaning protocol. Confirm agent, dilution, contact time and material compatibility. No pathogen-inactivation or sterilization claim is made. Do not mix chemicals or clean through electrical components.

Stop use if damage, leaks or electrical defects occur; qualified personnel must isolate and assess them. Order replacements by exact model and component details, not appearance.

Documentation and quote support

1-888-792-9315 · po@mymortuarycooler.com. Request current model-specific operating instructions, approved drawings and applicable warranty terms with your equipment schedule.

https://www.mymortuarycooler.com/pages/uspe-quote-request?brand=uspe-vet